





中国认可 国际互认 检测 TESTING CNAS L0599

Test Report SL52035273156201TX Date:July 23,2020 Page 1 of 4

SHANGHAI ASCENTET MEDICAL EQUIPMENT CO., LTD 4 / F, BUILDING 6, NO. 351, SIZHUAN ROAD, SONGJIANG DISTRICT, SHANGHAI

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)Surgical Gown

Buyer : Mr. Zhang
Composition : (A)45G SMS
Sample Color : (A)Blue

Style No. : Surgical Gown Order No. : AS-20200510

Supplier : SHANGHAI ASCENTET MEDICAL EQUIPMENT CO., LTD

Country of Origin : Canada

Sample Receiving Date : Jul 14, 2020

Testing Period : Jul 14, 2020 - Jul 23, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).

Test Performed : Selected test(s) as requested by applicant

Signed for and on behalf of SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Gan

Sara Guo (Account Executive)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available or request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-a-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerise to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report reter only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or, email: CN.Doccheck@sgs.com



Test Report

SL52035273156201TX

Date:July 23,2020

Page 2 of 4

Test Result

Liquid Barrier Performance and Classification of Protective Apparel and Drapes Intended for Use in **Health Care Facilities**

(ANSI/AAMI PB70:2012)

Section 4.2.1 Water Resistance: Impact Penetration Test

(AATCC 42-2017)

As received

Weight of blotter gained (g)	1#	2#	3#	4#	5#
Area A (Critical zone-front)	0.0	0.0	0.0	0.0	0.0
Area B (Critical zone-sleeve)	0.0	0.0	0.0	0.0	0.0
Area C(Front)	0.0	0.0	0.0	0.0	0.0
Area D(Back)	0.0	0.0	0.0	0.0	0.0
Seam between areas C&D	0.0	0.0	0.0	0.0	0.0

Remark:

- Level 1: all critical zone components shall have a blotter weight gain of no more than 4.5grams(g)
- Level 2: all critical zone components shall have a blotter weight gain of no more than 1.0 grams(g)
- Level 3: all critical zone components shall have a blotter weight gain of no more than 1.0 grams(g)

Section 4.2.1 Water Resistance: Hydrostatic Pressure Test

(AATCC 127-2018; Hydrostatic Head; Rate of increase of water pressure:60mbar/min; temp. of distilled water: 21℃, fabric face side of water)

As received

Water Column (cmH ₂ O)	1#	2#	3#	4#	5#
Area A (Critical zone-front)	>50.0	>50.0	>50.0	>50.0	>50.0
Area B (Critical zone-sleeve)	>50.0	>50.0	>50.0	>50.0	>50.0
Area C(Front)	>50.0	>50.0	>50.0	>50.0	>50.0
Area D(Back)	>50.0	>50.0	>50.0	>50.0	>50.0
Seam between areas C&D	>50.0	>50.0	>50.0	>50.0	>50.0

Remark:

- Level 2: all critical zone components shall have a hydrostatic resistance of at least 20cmH₂O
- Level 3: all critical zone components shall have a hydrostatic resistance of at least 50cmH₂O



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its Intervention only and within the limits of Client's Instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

3" Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233 中国・上海・徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 t (86-21) 61402666

f (86-21) 64958763

www.sgsgroup.com.cn e sgs.china@sgs.com



Test Report

SL52035273156201TX

Date: July 23,2020 Page 3 of 4

Barrier performance of each component and final classification commended

	Impact Penetration Test AATCC 42(g)	Hydrostatic Pressure Test AATCC 127(cmH2O)	Resistance to Bacteriophage Phi-X174 ASTM F 1671	Level	Final classification
Area A (Critical zone-front)	0.0	>50.0	/	Level 3	
Area B (Critical zone-sleeve)	0.0	>50.0	/	Level 3	
Area C(Front)	0.0	>50.0	/	Level 3	Level 3
Area D(Back)	0.0	>50.0	/	Level 3	
Seam between areas C&D	0.0	>50.0	/	Level 3	

Remark:

- The barrier performance of all critical zone components, including seams and points of attachments, shall be determined;
- The critical zone comprises at least area A and B. The classification of the surgical gown is based on the lower-performing component of the two;
- The back of the surgical gown (area D) may be nonprotective;
- Seams between protective and nonprotective areas have no barrier requirements;
- Seams between two protective areas are required to have at least the barrier performance of the lower-performing area;
- Level 1: Impact Penetration Test-AATCC 42: ≤4.5g;
- Level 2: Impact Penetration Test-AATCC 42: ≤1.0g; Hydrostatic Pressure Test-AATCC 127: ≥20cmH₂O;
- Level 3: Impact Penetration Test-AATCC 42: ≤1.0g; Hydrostatic Pressure Test-AATCC 127: ≥50cmH₂O;
- Level 4: Resistance to Bacteriophage Phi-X174-ASTM F 1671: Pass.
- Only test the barrier performance of the sample. Label request, construction and other classed in ANSI/AAMI PB70-2012 is not checked.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its Intervention only and within the limits of Client's Instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

3^{rr}Building,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国 • 上海 • 徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 t (86-21) 61402666

f (86-21) 64958763 f (86-21) 64958763 www.sgsgroup.com.cn e sgs.china@sgs.com




Test Report

SL52035273156201TX

Date:July 23,2020

Page 4 of 4



The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions/Terms

3^{rr}Building,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国 • 上海 • 徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f t (86-21) 61402666 f

f (86–21) 64958763 www.sgsgroup.com.cn f (86–21) 64958763 e sgs.china@sgs.com